

PRD 0-9 Conversion Chart

Level	Range
0	0...12 $\mu\text{R/h}$
1	12...27 $\mu\text{R/h}$
2	27...62 $\mu\text{R/h}$
3	62...142 $\mu\text{R/h}$
4	142...329 $\mu\text{R/h}$
5	329...760 $\mu\text{R/h}$
6	761 $\mu\text{R/h}$...1.75 mR/h
7	1.75...4.05 mR/h
8	4.05...9.4 mR/h
9	9.4...25 mR/h

0-2 – Normal Operating Conditions

3-6 – Be Alert – Determine source of radiation if possible. Notify Dispatch and start a PRD 1 Incident. If unable to determine source but can recreate the alarm – Notify Dispatch – PRD 2 Incident – Notify Shift Commander or Station 9 Officers for guidance.

7-9 – CAUTION – Retreat from area. Notify Dispatch – PRD 2 Response